



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/603,952	06/25/2003	Peter Lyon Harris	292-PDD-99-20-CON-[70P2]	3111
79990	7590	12/03/2010		
C. R. Bard, Inc. Bard Peripheral Vascular, Inc. 1415 W. 3rd Street P.O. Box 1740 Tempe, AZ 85280-1740			EXAMINER WILLSE, DAVID H	
			ART UNIT 3738	PAPER NUMBER
		NOTIFICATION DATE	DELIVERY MODE	
		12/03/2010	ELECTRONIC	

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

BPVIP.Docket@crbard.com
Jacki.Daspit@crbard.com

**Advisory Action
Before the Filing of an Appeal Brief**

Application No.	Applicant(s)	
10/603,952	HARRIS ET AL.	
Examiner	Art Unit	
David H. Willse	3738	

—The MAILING DATE of this communication appears on the cover sheet with the correspondence address —

THE REPLY FILED November 15, 2010, FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE.

1. The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods:

- a) The period for reply expires ____ months from the mailing date of the final rejection.
 b) The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.
 Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

NOTICE OF APPEAL

2. The Notice of Appeal was filed on _____. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a).

AMENDMENTS

3. The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because
 (a) They raise new issues that would require further consideration and/or search (see NOTE below);
 (b) They raise the issue of new matter (see NOTE below);
 (c) They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or
 (d) They present additional claims without canceling a corresponding number of finally rejected claims.

NOTE: _____. (See 37 CFR 1.116 and 41.33(a)).

4. The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324).

5. Applicant's reply has overcome the following rejection(s): _____.

6. Newly proposed or amended claim(s) _____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).

7. For purposes of appeal, the proposed amendment(s): a) will not be entered, or b) will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.

The status of the claim(s) is (or will be) as follows:

Claim(s) allowed: _____.

Claim(s) objected to: _____.

Claim(s) rejected: _____.

Claim(s) withdrawn from consideration: _____.

AFFIDAVIT OR OTHER EVIDENCE

8. The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e).

9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fail to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1).

10. The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached.

REQUEST FOR RECONSIDERATION/OTHER

11. The request for reconsideration has been considered but does NOT place the application in condition for allowance because:
 See Continuation Sheet

12. Note the attached Information Disclosure Statement(s). (PTO/SB/08) Paper No(s). _____

13. Other: _____.

/David H Willse/
 Primary Examiner, Art Unit 3738

Continuation of 11. does NOT place the application in condition for allowance because: The Applicant's remarks offer no evidence (such as technical dictionaries or bulletins or journal articles) to support the apparent allegation (Applicant's reply of November 15, 2010: paragraph bridging pages 2 and 3) that "uniform surface" (claim 1, line 3) implies (to a person of ordinary skill in the art) a constant or non-varying diameter along the longitudinal axis of a tubular portion or segment. Regarding Pintauro et al. and claim 1, leakage does not occur when the device is attached into a complementarily shaped oblique cut or when the device is entirely implanted within a vascular structure having an ostium (in a manner similar to that illustrated in Figure 13 of Pintauro et al.). Regarding Matterson and claim 27, the central axis extends beyond the oblique cut (even though the axis is at an angle to the cut), and the inner arc lengths decrease in an outward direction along this central axis (because of said cut). Other issues are adequately addressed in the final Office action of September 16, 2010.